

Table 1-1. Model 5342A Specifications

<b>INPUT CHARACTERISTICS</b>	
<p><b>INPUT 1:</b>  <b>Frequency range:</b> 500 MHz to 18 GHz  <b>Sensitivity:</b>            500 MHz to 12.4 GHz -25 dBm            12.4 GHz to 18 GHz -20 dBm  <b>Maximum input:</b> +5 dBm (see Options 002, 003 for higher level).  <b>Dynamic range:</b>            500 MHz to 12.4 GHz 30 dB            12.4 GHz to 18 GHz 25 dB  <b>Impedance:</b> 50 ohms, nominal  <b>Connector:</b> Precision Type N female  <b>Damage level:</b> +25 dBm, peak  <b>Coupling:</b> dc to load, ac to instrument.  <b>SWR:</b>            &lt;2:1, 500 MHz—10 GHz            &lt;3:1, 10 GHz—18 GHz  <b>FM tolerance:</b> Switch selectable (rear panel)  <b>FM (wide):</b> 50 MHz p-p worst case.  <b>CW (normal):</b> 20 MHz p-p worst case.            For modulation rates from dc to 10 MHz.  <b>AM tolerance:</b> Any modulation index provided the minimum signal level is not less than the sensitivity specification.  <b>Automatic amplitude discrimination:</b> Automatically measures the largest of all signals present, providing that signal is 6 dB above any signal within 500 MHz; 20 dB above any signal, 500 MHz—18 GHz.  <b>Modes of operation:</b>  <b>Automatic:</b> Counter automatically acquires and displays highest level signal within sensitivity range.  <b>Manual:</b> Center frequency entered to within <math>\pm 50</math> MHz to true value.  <b>Acquisition time:</b>  <b>Automatic mode:</b> Normal FM 530 ms worst case; wide FM 2.4 s worst case.  <b>Manual mode:</b> 80 ms after frequency entered.</p>	<p><b>External time base:</b> Requires 10 MHz, 2.0V peak-to-peak sine wave or square wave into 1 K<math>\Omega</math> via rear panel BNC connector. Switch selects either internal or external time base.</p> <p style="text-align: center;"><b>OPTIONAL TIME BASE (OPTION 001)</b></p> <p>Option 001 provides an oven-controlled crystal oscillator time base, 10544A (see separate data sheet), that results in better accuracy and longer periods between calibration.  <b>Crystal frequency:</b> 10 MHz  <b>Stability:</b>  <b>Aging rate:</b> <math>&lt;5 \times 10^{-10}</math>/day after 24-hour warm-up.  <b>Temperature:</b> <math>&lt;7 \times 10^{-9}</math> over the range 0°C to 50°C.  <b>Short term:</b> <math>&lt;1 \times 10^{-11}</math> for 1 second average time.  <b>Line variation:</b> <math>&lt;1 \times 10^{-10}</math> for 10% change from nominal.  <b>Warm-up:</b> <math>&lt;5 \times 10^{-9}</math> of final value 20 minutes after turn-on, at 25°C.</p> <p style="text-align: center;"><b>AMPLITUDE MEASUREMENT (OPTION 002)</b></p> <p>Option 002 provides the capability of measuring the amplitude of the incoming sine wave signal, and simultaneously displaying its frequency (MHz) and level (dBm). The maximum operating level and the top end of dynamic range are increased to +20 dBm. Amplitude offset to 0.1 dB resolution may be selected from front panel pushbuttons.</p>
<p><b>INPUT 2:</b>  <b>Frequency range:</b> 10 Hz to 520 MHz Direct Count.  <b>Sensitivity:</b>            50<math>\Omega</math> 10 Hz to 520 MHz 25 mV rms.            1 M<math>\Omega</math> 10 Hz to 25 MHz 50 mV rms.  <b>Impedance:</b> Selectable: 1 M<math>\Omega</math>, &lt;50 pF or 50<math>\Omega</math> nominal.  <b>Coupling:</b> ac  <b>Connector:</b> Type BNC female.  <b>Maximum input:</b> 50<math>\Omega</math> 3.5V rms (+24 dBm) or 5V dc fuse protected; 1 M<math>\Omega</math> 200V dc +5.0V rms.</p> <p style="text-align: center;"><b>TIME BASE</b></p> <p><b>Crystal frequency:</b> 10 MHz  <b>Stability:</b>  <b>Aging rate:</b> <math>&lt;1 \times 10^{-7}</math> per month.  <b>Short term:</b> <math>&lt;1 \times 10^{-9}</math> for 1 second average time.  <b>Temperature:</b> <math>&lt;\pm 1 \times 10^{-6}</math> over the range 0°C to 50°C.  <b>Line variation:</b> <math>&lt;\pm 1 \times 10^{-7}</math> for 10% change from nominal.  <b>Output frequency:</b> 10 MHz <math>\geq 2.4</math>V square wave (TTL compatible); 1.5V peak-to-peak into 50<math>\Omega</math> available from rear panel BNC.</p>	<p><b>INPUT 1:</b>  <b>Frequency range:</b> 500 MHz—18 GHz.  <b>Dynamic range (frequency and level):</b>            -22 dBm to +20 dBm 500 MHz to 12.4 GHz            -15 dBm to +20 dBm 12.4 GHz to 18 GHz  <b>Maximum operating level:</b> +20 dBm  <b>Damage level:</b> +25 dBm, peak  <b>Resolution:</b> 0.1 dB  <b>Accuracy:</b> <math>\pm 1.5</math> dB (excluding mismatch uncertainty).  <b>SWR:</b>            &lt;2:1 (amplitude measurement).            &lt;5:1 (frequency measurement).  <b>Measurement time:</b> 100 ms + frequency measurement time.  <b>Display:</b> Simultaneously displays frequency to 1 MHz resolution and input level. (Option 011 provides full frequency resolution on HP-IB output.)</p> <p><b>INPUT 2: (50<math>\Omega</math> impedance only)</b>  <b>Frequency range:</b> 10 MHz—520 MHz  <b>Dynamic range (frequency and level):</b>            -17 dBm to +20 dBm.  <b>Damage level:</b> +24 dBm, peak  <b>Resolution:</b> 0.1 dBm.  <b>Accuracy:</b> 1.5 dB (excluding mismatch uncertainty).  <b>SWR:</b> &lt;1.8:1  <b>Measurement time:</b> 100 ms + frequency measurement time.  <b>Display:</b> Simultaneously displays frequency to 1 MHz resolution and input level.</p>

Table 1-1. Model 5342A Specifications (Continued)

EXTENDED DYNAMIC RANGE (OPTION 003)	GENERAL
Option 003 provides an attenuator that automatically extends the dynamic range of operation for input 1.	<b>Accuracy:</b> $\pm 1$ count $\pm$ time base error.
<b>INPUT 1:</b>	<b>Resolution:</b> Front panel pushbuttons select 1 Hz to 1 MHz.
<b>Frequency range:</b> 500 MHz to 18 GHz	<b>Residual stability:</b> When counter and source use common time base or counter uses external higher stability time base, $< 4 \times 10^{-11}$ rms typical.
<b>Sensitivity:</b>	<b>Display:</b> 11-digit LED display, sectionalized to read GHz, MHz, kHz, and Hz.
500 MHz to 12.4 GHz -22 dB	<b>Self-check:</b> Selected from front panel pushbuttons. Measures 75 MHz for resolution chosen.
12.4 GHz to 18 GHz -15 dBm	<b>Frequency offset:</b> Selected from front panel pushbuttons. Displayed frequency is offset by entered value to 1 Hz resolution.
<b>Maximum operating level:</b> +20 dBm.	<b>Sample rate:</b> Variable from less than 20 ms between measurements to HOLD which holds display indefinitely.
<b>Dynamic range:</b>	<b>IF out:</b> Rear panel BNC connector provides 25 MHz to 125 MHz output of down-converted microwave signal.
500 MHz to 12.4 GHz 42 dB	<b>Operating temperature:</b> 0°C to 50°C.
12.4 GHz to 18 GHz 35 dB	<b>Power requirements:</b> 100/120/220/240V rms, +5%, -10%, 48—66 Hz; 100 VA max.
<b>Damage level:</b> +25 dBm, peak	<b>Accessories furnished:</b> Power cord, 229 cm (7½ ft.)
<b>SWR:</b> <5:1	<b>Size:</b> 133 mm H $\times$ 213 mm W $\times$ 498 mm D (5¼" $\times$ 8¾" $\times$ 19⅝")
<b>DIGITAL-TO-ANALOG CONVERTER (OPTION 004)</b>	<b>Weight:</b> Net 9.1 kg (20 lbs.). Shipping 12.7 kg (28 lbs.).
Option 004 provides the ability to convert any three consecutive displayed digits into an analog voltage output. A display of 000 produces 0V output; 999 produces 9.99V full scale.	
<b>Accuracy:</b> $\pm 5$ mV, $\pm 0.3$ mV/°C (from 25°C)	
<b>Conversion Speed:</b> $< 50 \mu s$ to $\pm 0.01\%$ of full scale reading.	
<b>Resolution:</b> 10 mV	
<b>Output:</b> 5 mA. Impedance $< 1.0$ ohm.	
<b>Connector:</b> Type BNC female on rear panel.	

## 1-5. SAFETY CONSIDERATIONS

1-6. This product is a Safety Class I instrument (provided with a protective earth terminal). Safety information pertinent to the operation and servicing of this instrument is included in appropriate sections of this manual.

## 1-7. INSTRUMENT IDENTIFICATION

1-8. Hewlett-Packard instruments have a 2-section, 10-character serial number (0000A00000), which is located on the rear panel. The four-digit serial prefix identifies instrument changes. If the serial prefix of your instrument differs from that listed on the title page of this manual, there are differences between this manual and your instrument. Instruments having higher serial prefixes are covered with a "Manual Changes" sheet included with this manual. If the change sheet is missing, contact the nearest Hewlett-Packard Sales and Service Office listed at the back of this manual. Instruments having a lower serial prefix than that listed on the title page, are covered in Section VII.

## 1-9. ACCESSORIES

1-10. Table 1-2 lists accessory equipment supplied and Table 1-3 lists accessories available.

Table 1-2. Equipment Supplied

DESCRIPTION	HP PART NUMBER
Detachable Power Cord 229 cm (7½ feet long)	8120-1378